



महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड.

जिल्हा सामान्य रुग्णालय परिसर, बीच रोड, लिमये वाडी, अलिबाग, जि. रायगड - ४०२ २०१

GOVERNMENT MEDICAL COLLEGE, ALIBAG - RAIGAD

DISTRICT CIVIL HOSPITAL AREA, ALIBAG BEACH ROAD, LIMYEWADI, ALIBAG DIST. RAIGAD - 402 201

Email- deangmcalibag@gmail.com

Phone No. 02141-299214

www.gmchalibag.in/gmch

जा. क्र. शा. वै. म. अ. / Ophthalmology / दरपत्रक मागणी / 2824 / 2023

दिनांक. २७ . १२ . २०२३

दरपत्रक मागणी पत्र.

प्रति,

-----,
-----,
-----.

विषय - नेत्रशल्यचिकिस्ताशास्त्र विभागा करीता नमुद साधन/यंत्रसामुग्रीसाठी दरपत्रके मोहोरबंद लिफाफे सादर करणेबाबत.

शासकीय वैद्यकीय महाविद्यालय, अलिबाग - रायगड, नेत्रशल्यचिकिस्ताशास्त्र विभागसाठी नमुद साधन/यंत्रसामुग्री ची आवश्यकता आहे, त्या अनुषंगाने नेत्रशल्यचिकिस्ताशास्त्र विभागा करीता नमुद साधन/यंत्रसामुग्री ची खरेदी ही खालील नमुन्यात मोहोरबंद लिफाफा विहित मुदतीत या कार्यालयाच्या आवक शाखेत खालील अटी व शर्ती नुसार सादर करावेत.

Sr. No	Product	Specification	Price Including all Taxes
1	Auto Refractometer with Kerotometer	Attached	
2	Wet Field Electric Cautery	Attached	
3	Snellens Chart with remote control	Attached	
4	Color Vision - Original Ishihara	Attached	
5	Schiotz's Tonometer	Attached	
	Total Including All Taxes		

अ. नं.

- दरपत्रक मोहोरबंद लिफाफा, दि. ०२ जानेवारी २०२४ रोजी सांयकाळी ०५.०० वाजेपर्यंत कार्यालयास पोहचतील या बेताने पाठवावे. जाहीर केल्याला तारखेनंतर अथवा कार्यालयीन वेळेनंतर आलेले दरपत्रक विचारत घेतले जाणार नाही / ग्राह्य धरले जाणार नाही.
- दरपत्रक मोहोरबंद लिफाफा अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड यांचे नावे विहित मुदतीत कार्यालयास सादर करावा, त्याचप्रमाणे दरपत्रक मोहोरबंद लिफाफयावर "नेत्रशल्यचिकिस्ताशास्त्र विभागा" करीता दरपत्रक लिहून सादर करावे.
- लिफाफयावर डाव्या बाजूस दरपत्रक, लिफाफा जमा करण्याचा अंतिम दिनांक, दरपत्रक धारकाच्या नावाचा शिक्का इ. नमुद असणे आवश्यक आहे.

४. दरपत्रक ग्राहय कालावधी हा पुरवठा आदेश नोंदविलेल्या दिनांका पासुन १२ महिन्यांचा राहिल.
५. प्रस्तावित उदिदष्टांकरीता दरपत्रक आकारण्यात आलेले दर हे बाजारभावा पेक्षा अधिक नसावेत. सर्वात कमीतकमी दराचे दरपत्रकाचा विचार करण्यात येईल.
६. सदर उदिदष्टांची किंमत, सर्व करांसहीत एकुण किंमत अशी दर्शविण्यात यावी. दरपत्रकात खाडाखोड करण्यात येवू नये.
७. दरपत्रक मोहोरबंद लिफाफयात सादर करावायाचे दस्ताऐवज (Documents)- अधिकृत लेटर हेडवर दरपत्रकसादर करावे. १) दुकानाचा परवाना, २) व्यवसाय नोंदणी प्रमाणपत्र, ३) जी.एस.टी.क्रमांक पत्र.
८. सदर उपकरण इतर कोणत्या ही शासकीय संस्थेस पुरवठा केल्याचे प्रमाणपत्र असल्यास प्राधान्य.
९. ज्या पात्र दरपत्रक धारकाचे दर हे न्यूनतम असतील अशा यशस्वी दरपत्रक धारकास आवश्यकते नुसार नुमद चा पुरवठा करण्याचे आदेश देण्यात येतील.
१०. अपवादात्मक परिस्थितीमुळे उक्त कोणत्याही अटीत आवश्यकतेनुसार अंशतः बदल करणे अथवा प्राप्त झालेली दरपत्रके कोणतेही कारण न देता स्वीकारणे / नाकारणे तसेच कोणतही स्तरावर खरेदी प्रक्रिया रददकरणे इ. अधिकारी संस्थाप्रमुख (अधिष्ठाता) यांनी राखून ठेवले आहेत.
११. काम पुर्ण झाल्यानंतर तीन प्रतीत देयकेही अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड यांचे नावाने देण्यात यावेत.
१२. देयकाची अदायगी ही शासनाकडून अनुदान प्राप्त होण्याच्या अधीन राहून करण्यात येईल. देयक अदा करण्यास विलंब झाल्यास त्यावर कोणत्याही प्रकारचे विलंब शुक्ल अदा केले जाणार नाही.
१३. वरील नमुद वस्तुंची नग संख्या ही मा. अधिष्ठाता यांच्या आदेशान्वये / प्रशासकीय उपलब्धतेनुसार कमी किंवा वाढवली जाऊ शकते.
१४. पुरवठाधारकाचे पुर्ण नाव, भ्रमणध्वनी आणि संपुर्ण पत्ता, संकेतांक क्र. (Email address along with Website) हा Envelope वर नमुद असणे अर्निवार्य आहे.

PP-9

अधिष्ठाता,
शासकीय वैद्यकीय महाविद्यालय,
अलिबाग-रायगड.

DEAN
Government Medical Collage

प्रत - शासकीय वैद्यकीय महाविद्यालय, तसेच जिल्हा सामान्य रुग्णालय, अलिबाग-रायगड यांच्या सुचना फलकावर दरपत्रक मागविण्या संदर्भात दिनांक २७ डिसेंबर २०२३ सुचना देण्यात आल्या आहेत.

Schiotz's Tonometer

Schiotz's Tonometer

- Autoclavable
- Perfect reading of the scale with red pointer
- Complete in a black deluxe case with velvet-look inserts
- High-quality agate bearing for an extremely long service life
- Supplied with three weights (5.5 g, 7.5 g, 10 g) and a conversion table
- All vital parts are made of stainless steel, other parts are chromeplated
- Precision measurement on a scale of 0 to 20 subdivisions and 0 to -1 sub-divisions

Specification 5 (polished weights with inspection certificate)

- schiötz A, with straight scale
- schiötz B, with inclined scale **Specification 3**
- schiötz C, with inclined scale

USA FDA and/or EUROPIAN CE and/or Country of origin CE

Sr. no	Equipment	Specifications
1	Snellens chart with or without remote control	<p style="text-align: center;">Low End</p> <ol style="list-style-type: none"> 1. Rotating drum with visual acuity from 6/60 to 6/6 2. All National languages inclusive eg – English, Marathi, Hindi, E Chart 3. Wall mounted <p style="text-align: center;">Medium End / High End (Price Range – 30000-40000)</p> <ol style="list-style-type: none"> 1. LED Screen with visual acuity from 6/60 to 6/6 2. All National languages inclusive eg – English, Marathi, Hindi 3. Landolt's E chart 4. Picture chart paediatric 5. Wall mounted screen – 15 inch and above 6. Color vision chart 7. ETDRS chart 8. Log mar unit conversion available 9. Operating remote with battery – 1 additional remote 10. Worth four dot test 11. Adjustable to distance 12. US FDA/ European CE/ Country of origin CE

11/21/2020


Technical Specification
Auto Refractometer with kerotometer

OBJECTIVE REFRACTMETER MODE : Sphere range : - 25 D to +22D (step: $\pm 0.12D \pm 0.25 D$) Cylinder Range : 0 D to $\pm 10 D$ (step : $\pm 0.12D / \pm 0.25 D$) Axis range : 0 degree to 180 degree (step : 1 degree)
Corneal curvature mode : Refraction index : 1.3375 Corneal curvature radius : 5 to 10 mm (Step 0.01 mm) Corneal refraction : 67.5 D to 33.75 D (Step $\pm 0.12D / \pm 0.25D$) Corneal astigmatism : 0 D to $\pm 10 D$ (step : $\pm 0.12 D / \pm 0.25 D$) Corneal astigmatism Axis range : 0 degree to 180 degree (step : 1 degree)
Papillary distance measurement 20-85 mm in 1 mm step
Output : RS232C/USB 1.1
Dimension (WxDxH) 288 x 509 x 432 - 462 mm
Weight approx 20 kg
Power supply 100- 240 V AC 50/60 Hz
Contact lenses base curve
Voltage stabilizer Must
CE/US-FDA/European CE

Signature

Wet Field Electric Cautey

		<p>1. Technical Specifications</p> <ul style="list-style-type: none">• Should have use solid –state technology.• Should have LED indicator for power output.• Should have foot control.• Should have provision to connect wide range of bipolar forceps and haemostatic eraser to facilitate most ophthalmic surgery procedures.• Should be supplied with 2 sets each of reusable, autoclavable cord for bipolar forceps and hemostatic erasers.• Should be supplied with 2 standard ophthalmic bipolar forceps.• Should be supplied with 2 bipolar eraser with 18G tip.• Should work with input 200 to 240V 50Hz supply.
		<p>2.</p> <ul style="list-style-type: none">• Should be FDA or CE or BIS or ISO and STQC certified Product.
		<p>3.</p> <ul style="list-style-type: none">• User Manual and Service manual in English must be provided.
		<p>4.</p> <ul style="list-style-type: none">• The equipment and all accessories should be installed, tested and commissioned at the Department of ophthalmology, JIPMER, Pondicherry free of cost.
		<p>5.</p> <ul style="list-style-type: none">• The Equipment including all accessories including bought out items should be under Warranty for a period of 3 Years after successful Commissioning.• Annual maintenance contract rates for 2 years after warranty must be quoted. All spare parts and consumables should be available with supplier or principals for a period of at least 10 years.• Must have after sales service facility in south India preferably in Pondicherry or in Tamilnadu.
		<p>6. Other tender conditions</p> <ul style="list-style-type: none">• Suppliers should have been in the market for at least 3 years.• Suppliers should have made a large number of installations, within the last five years in the country in reputed institutions and preferably in Government Hospitals with Proven track record of excellent after sales support for this system• User-list of at least 5 users must be enclosed.



x/ol on 500

**Color vision - Original
Ishihara**

5 Color vision - Original Ishihara	1. All 38 plates present 2. Hard bound book 3. Interpretation chart for color blindness
---------------------------------------	---

~~11/12/23~~

costly

ku. 20, 7 20/-
Original - Kanekole version.

2000/- on E. Upkaran
but Quality not standard.

~~mark~~