



महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड.
जिल्हा सामान्य रुग्णालय परिसर, बीच रोड, अलिबाग. जि.रायगड

Email- deangmcalibag@gmail.com

Phone No. 02141-29921

www.gmcalibag.in/gmch

जा.क्र.शावैमअ/ शुद्धिपत्रक/ विकृतीशास्त्र/ 308 / 2028

दि. १५/०२/२०२४

शुद्धिपत्रक:

विषय :- विकृतीशास्त्र विभागासाठी दरपत्रके मोहोरबंद लिफाफेची मुदतवाढ देणेबाबत.


संदर्भ :- जा.क्र.शावैमअ/दरपत्रक/ विकृतीशास्त्र/३३६/२०२४ दिनांक.०८/०२/२०२४.

अलिबाग येथे १०० विद्यार्थी प्रवेश क्षमतेचे शासकीय वैद्यकीय महाविद्यालय सन २०२१-२२ या शैक्षणिक वर्षापासून सुरु करण्यास शासनाने व राष्ट्रीय आर्विज्ञान आयोग नवी दिल्ली, यांनी मान्यता प्रदान केलेली आहे. त्या अनुषंगाने महाविद्यालयातील विकृतीशास्त्र विभागाला व्याख्याने व प्रात्यक्षिक घेण्याकरिता खालील प्रमाणे अतिरिक्त यंत्रसामुग्रीची आवश्यकता आहे. त्या अनुषंगाने खालील नमुन्यात मोहोरबंद लिफाफा विहित मुदतीत या कार्यालयाच्या आवक शाखेत खालील अटी व शर्ती नुसार सादर करण्याकरीता दिनांक ०८/०२/२०२४ रोजी आदेशीत केले होते परंतु दिनांक १५/०२/२०२४ रोजी ५ वाजेपर्यंत कोणत्याही पुरवठाधारकाचे दरपत्रके कार्यालयास प्राप्त झालेले नसल्यामुळे अतिरिक्त सात दिवसांची मुदतवाढ देण्या संदर्भात कार्यवाही करण्यात येत आहे.

Sr. No.	Particulars	Specification	Quantity	Unit Price
1	Water Bath	Temperature range – RT to 90degree Celsius and maintained automatically with Accuracy ± 1 °C Black water tray with built-in LED backlight OLED touch control panel and slide dryer Heating temperature: up to 90°C Dimensions (WxDxH): 280 mm x 280 mm x 105 mm Water tray dimensions (WxDxH): 230 mm x 180 mm x 53 mm Heating speed: 20 mins +/- 10%, (220V Voltage), 8 lit.	1	Including all taxes
2	Refrigerator	Capacity 244 litres, Freezer Capacity 50 Litres, Cooling Retention upto 24 hours, Toughened glass, 4 to 5 shelves Bright LED Prism Lens, Voltage: 230 Volts, Defrost System Automatic, Item Weight less than 55 kg	1	
3	Scalpel blade	Disposable Surgical Blade no.10 Tool Scalpel Blades, High Carbon Steel Blades. Individually Wrapped no.10 Surgical Blades	10 box (100pc)	
4	Scalpel handle	Stainless Steel Scalpel Handle for surgical blade Handle No (Size-3)	10	
5	L Mould1	Specimen Holder Used for holding the wax block for taking sections on the microtome. Sizes :- 37 x 25 x 15 mm pair,	10	
6	L Mould2	Specimen Holder Used for holding the wax block for taking sections on the microtome 50 x 25 x 15 mm pair,	10	

7	L Mould3	Specimen Holder Used for holding the wax block for taking sections on the microtome. 75 × 30 × 19 mm pair	10
8	Tissue cassette	Stainless steel tissue cassettes size 3 × 2.5 × 0.4 cm.	100
9	Centrifuge (heavy duty)	Type of product: Laboratory centrifuges; Supply: 220- 240 Volts 50 Hz single phase Capacity (ml): 400; Rotor capacity (ml): 16X15ML Type of head: Swing out head; No. of tubes: minimum 16 Max. speed (RPM): 5250; Max. RCF: 3600; Digital timer range: 0-60 minutes Warranty 1 yr	2
10	Centrifuge (light weight)	Type of product: Laboratory centrifuges; Supply: 220 - 240 Volts 50 Hz single phase Rotor capacity (ml): 8x15 ml; Type of head: Angle rotor head No. of tubes: 8; Max. speed (RPM): 3500 Max. RCF: 1450; Weight: 6 kg; W x D x M: 210 x 210 x 250 Warranty-1yr	8
11	Digital Weighing scale	Mounting Type Table Top and Bench Type Display Type 7 Segment LED display Power Supply 230V AC, 50Hz Dimensions 304(L) x 295(D) x 115(H) mm. +-0. 5 mm. Display Size 0.56" [14.22 mm.] Dual Display In-built Battery 6V / 4Ah Rechargeable Power Consumption 6W Display Digit 6 Digit Number of Keys 5 Keys Platform: Large platter with weighing capacity upto 50kg Warranty 1yr	4
12	Microtome blade – High profile	Stainless steel high profile blade-Compatible with Rotary Microtome HM 355S	1 box (50pc)
13	Microtome blade – Low profile	Stainless steel low profile blade for Rotary Microtome HM 355S	1 box (50pc)
14	Hot Plate	Technical Specifications --- Voltage - 230 V.a.c./50 Hz Power Consumption - 200 W. to 850 W. Ampere - 4 AMP Temperature range=50°C to 350°C	1
15	Tissue floatation bath	Dimensions: 15.5" Long x 14.5" Wide x 4.5" High Material: Stainless Steel Construction with Baked Gray Enamel Finish. Removable 9" x 5" x 2" deep glass dish (capacity 2 liters) Digital Temperature Display Ambient to 60 Degrees Celsius LED Control Panel Illuminated Black Background	1
16	Autoclave	Autoclave Portable is Single Walled. • Capacity 10L. Made of Stainless Steel 304 Q. Heavy Gauge. • The Gasket is made of Superior quality Silicon rubber which is easily Replaceable. • Electrically operated supplied with pressure gauge, safety valve cum steam release valve. TECHNICAL DATA • Working pressure upto 15 PSI Power Supply : 230 Volts A.C. 50 Hz Single Phase OPTIONAL ACCESSORIES •	1

		Stainless Steel Basket• Digital Temperature Controller & mer		
17	Hot Air Oven	Dry Hot Air Oven (50 liters) with temperature setting from RT + 10°C up to +300°C Internal dimension (W — D — H mm) : 420—395—350 Chamber volume (Litres) : 50 Temperature range : RT+10°~300°C Power consumption : 1100W Electrical requirements : 220V, 50Hz, Single phase	1	
18	Tissue embedding moulds1	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 16 x 16 x 6mm	1 box (12pc)	
19	Tissue embedding moulds2	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 22 x 22 x 6mm	1 box (12pc)	
20	Tissue embedding moulds3	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 32 x 25 x 6mm	1 box (12pc)	
21	Tissue embedding moulds4	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 38 x 25 x 6mm	1 box (12pc)	
22	Tissue embedding moulds5	Stainless Steel Tissue Embedding Moulds Biopsy Cryomolds 10 x 10 x 5mm	1 box (100pc)	
23	Surgical Scissor	Stainless steel 8 inch long blunt end, with sharp cutting	10	
24	Grossing knife	Large Grossing Knife, Blade Length=14 inches, sharp Premium grade stainless steel with plastic handle	10	



अधिष्ठाता,

शासकीय वैद्यकीय महाविद्यालय,

अलिबाग-रायगड



महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड.
जिल्हा सामान्य रुग्णालय परिसर, बीच रोड, अलिबाग. जि.रायगड

Email- deangmcalibag@gmail.com

Phone No.02141-29921

www.gmchalibag.in/gmch

जा.क्र.शावैमअ/दरपत्रक/ विकृतीशास्त्र/ २३६ /२०२४

दि. ०८/०२/२०२४

दरपत्रक मागणी पत्र.

प्रति,

विषय :- विकृतीशास्त्र विभागासाठी दरपत्रके मोहोरबंद लिफाफे सादर करणेबाबत.

अलिबाग येथे १०० विद्यार्थी प्रवेश क्षमतेचे शासकीय वैद्यकीय महाविद्यालय सन २०२१-२२ या शैक्षणिक वर्षापासून सुरु करण्यास शासनाने व राष्ट्रीय अर्विज्ञान आयोग नवी दिल्ली, यांनी मान्यता प्रदान केलेली आहे. त्या अनुषंगाने महाविद्यालयातील विकृतीशास्त्र विभागाला व्याख्यान व प्रात्यक्षिक घेण्याकरिता खालील प्रमाणे अतिरिक्त यंत्रसामुग्रीची आवश्यकता आहे. त्या अनुषंगाने खालील नमुन्यात मोहोरबंद लिफाफा विहित मुदतीत या कार्यालयाच्या आवक शाखेत खालील अटी व शर्ती नुसार सादर करावेत.

Sr. No.	Particulars	Specification	Quantity	Unit Price
1	Water Bath	Temperature range – RT to 90degree Celsius and maintained automatically with Accuracy ± 1 °C Black water tray with built-in LED backlight OLED touch control panel and slide dryer Heating temperature: up to 90°C Dimensions (WxDxH): 280 mm x 280 mm x 105 mm Water tray dimensions (WxDxH): 230 mm x 180 mm x 53 mm Heating speed: 20 mins +/- 10%, (220V Voltage), 8 lit.	1	Including all taxes
2	Refrigerator	Capacity 244 litres, Freezer Capacity 50 Litres, Cooling Retention upto 24 hours, Toughened glass, 4 to 5 shelves Bright LED Prism Lens, Voltage: 230 Volts, Defrost System Automatic, Item Weight less than 55 kg	1	
3	Scalpel blade	Disposable Surgical Blade no.10 Tool Scalpel Blades, High Carbon Steel Blades. Individually Wrapped no.10 Surgical Blades	10 box (100pc)	
4	Scalpel handle	Stainless Steel Scalpel Handle for surgical blade Handle No (Size-3)	10	
5	L Mould1	Specimen Holder Used for holding the wax block for taking sections on the microtome. Sizes :- 37 x 25 x 15 mm pair,	10	
6	L Mould2	Specimen Holder Used for holding the wax block for taking sections on the microtome 50 x 25 x 15 mm pair,	10	
7	L Mould3	Specimen Holder Used for holding the wax block for taking sections on the microtome.	10	

		75 × 30 × 19 mm pair	
8	Tissue cassette	Stainless steel tissue cassettes size 3 × 2.5 × 0.4 cm.	100
9	Centrifuge (heavy duty)	Type of product: Laboratory centrifuges; Supply: 220- 240 Volts 50 Hz single phase Capacity (ml): 400; Rotor capacity (ml): 16X15ML Type of head: Swing out head; No. of tubes: minimum 16 Max. speed (RPM): 5250; Max. RCF: 3600; Digital timer range: 0-60 minutes Warranty 1 yr	2
10	Centrifuge (light weight)	Type of product: Laboratory centrifuges; Supply: 220 - 240 Volts 50 Hz single phase Rotor capacity (ml): 8x15 ml; Type of head: Angle rotor head No. of tubes: 8; Max. speed (RPM): 3500 Max. RCF: 1450; Weight: 6 kg; W x D x M: 210 x 210 x 250 Warranty-1yr	8
11	Digital Weighing scale	Mounting Type Table Top and Bench Type Display Type 7 Segment LED display Power Supply 230V AC, 50Hz Dimensions 304(L) x 295(D) x 115(H) mm. +-0. 5 mm. Display Size 0.56" [14.22 mm.] Dual Display In-built Battery 6V / 4Ah Rechargeable Power Consumption 6W Display Digit 6 Digit Number of Keys 5 Keys Platform: Large platter with weighing capacity upto 50kg Warranty 1yr	4
12	Microtome blade – High profile	Stainless steel high profile blade-Compatible with Rotary Microtome HM 355S	1 box (50pc)
13	Microtome blade – Low profile	Stainless steel low profile blade for Rotary Microtome HM 355S	1 box (50pc)
14	Hot Plate	Technical Specifications --- Voltage - 230 V.a.c./50 Hz Power Consumption - 200 W. to 850 W. Ampere - 4 AMP Temperature range=50°C to 350°C	1
15	Tissue floatation bath	Dimensions: 15.5" Long x 14.5" Wide x 4.5" High Material: Stainless Steel Construction with Baked Gray Enamel Finish. Removable 9" x 5" x 2" deep glass dish (capacity 2 liters) Digital Temperature Display Ambient to 60 Degrees Celsius LED Control Panel Illuminated Black Background	1
16	Autoclave	Autoclave Portable is Single Walled. • Capacity 10L. Made of Stainless Steel 304 Q. Heavy Gauge. • The Gasket is made of Superior quality Silicon rubber which is easily Replaceable. • Electrically operated supplied with pressure gauge, safety valve cum steam release valve. TECHNICAL DATA • Working pressure upto 15 PSI Power Supply : 230 Volts A.C. 50 Hz Single Phase OPTIONAL ACCESSORIES • Stainless Steel Basket • Digital Temperature Controller & mer	1

17	Hot Air Oven	Dry Hot Air Oven (50 liters) with temperature setting from RT + 10°C up to +300°C Internal dimension (W — D — H mm) : 420—395—350 Chamber volume (Litres) : 50 Temperature range : RT+10°~300°C Power consumption : 1100W Electrical requirements : 220V, 50Hz, Single phase	1
18	Tissue embedding moulds1	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 16 x 16 x 6mm	1 box (12pc)
19	Tissue embedding moulds2	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 22 x 22 x 6mm	1 box (12pc)
20	Tissue embedding moulds3	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 32 x 25 x 6mm	1 box (12pc)
21	Tissue embedding moulds4	Stainless Steel Tissue Embedding Moulds Base Mould for Embedding Ring 38 x 25 x 6mm	1 box (12pc)
22	Tissue embedding moulds5	Stainless Steel Tissue Embedding Moulds Biopsy Cryomolds 10 x 10 x 5mm	1 box (100pc)
23	Surgical Scissor	Stainless steel 8 inch long blunt end, with sharp cutting	10
24	Grossing knife	Large Grossing Knife, Blade Length=14 inches, sharp Premium grade stainless steel with plastic handle	10

अ.नं.

१. दरपत्रक मोहरबंद लिफाफा, दि. १५/०२/२०२४ रोजी ०५:०० वाजेपर्यंत कार्यालयास पोहचतील या बेताने पाठवावे. जाहीर केलेल्या तारखेनंतर अथवा कार्यालयीन वेळेनंतर आलेले दरपत्रक विचारत घेतले जाणार नाही / ग्राहय धरले जाणार नाही.
२. दरपत्रक मोहोरबंद लिफाफा अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग. जि. रायगड यांचे नावे विहित मुदतीत कार्यालयास सादर करावा. त्याचप्रमाणे दरपत्रक मोहरबंद लिफाफयावर विकृतीशास्त्र पुरवठा विभागा करीता दरपत्रक लिहून सादर करावे.
३. लिफाफयावर डाव्या बाजूस दरपत्रक, लिफाफा जमा करण्याचा अंतिम दिनांक, दरपत्रक धारकाच्या नावाचा शिक्का इ. नमुद असणे आवश्यक आहे.
४. दरपत्रक ग्राहय कालावधी हा पुरवठा आदेश नोंदविलेल्या दिनांका पासुन बारा (१२) महिन्यांचा राहिल.
५. प्रस्तावित उदिदष्टांकरिता दरपत्रक आकारण्यात आलेले दर हे बाजारभावापेक्षा अधिक नसावेत. सर्वात कमीतकमी दराचे दरपत्रकाचा विचार करण्यात येईल.
६. सदर उदिदष्टांची किंमत, सर्व करांसहीत एकुण किंमत अशी दर्शविण्यात यावी. दरपत्रकात खाडाखोड करण्यात येवू नये.
७. दरपत्रक माहोरबंद लिफाफयात सादर करावयाचे दस्ताऐवज (Documents) अधिकृत लेटर हेडवर दरपत्रक सादर करावे. १) दुकानाचा परवाना, २) व्यवसाय नोंदणी प्रमाणपत्र, ३) जी.एस.टी. क्रमांक पत्र.
८. सदर उपकरण इतर कोणत्या ही शासकीय संस्थेस पुरवठा केल्याचे प्रमाणपत्र असल्यास प्राधान्य.
९. ज्या पात्र दरपत्रक धारकाचे दर हे न्यूनतम असतील अशा यशस्वी दरपत्रक धारकास आवश्यकते नुसार विकृतीशास्त्र विभागासाठीचा पुरवठा करण्याचे आदेश देण्यात येतील.
१०. अपवादात्मक परिस्थितीमुळे उक्त कोणत्याही अटीत आवश्यकतेनुसार अंशतः बदल करणे अथवा प्राप्त झालेली दरपत्रके कोणतेही कारण न देता स्वीकारणे/नाकारणे तसेच कोणतेही स्तरावर खरेदी प्रक्रिया रद्द करणे इ. अधिकारी संस्थाप्रमुख (अधिष्ठाता) यांनी राखून ठेवले आहेत.

११. काम पुर्ण झाल्यानंतर तीन प्रतीत देयकेही अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग जि. रायगड यांचे नावाने देण्यात यावेत.
१२. देयकाची अदायगी ही शासनाकडून अनुदान प्राप्त होण्याच्या अधीन राहून करण्यात येईल. देयक अदा करण्यास विलंब झाल्यास त्यावर कोणत्याही प्रकारचे विलंब शुक्ल अदा केले जाणार नाही.
१३. वरील नमुद वस्तुंची नग संख्या ही मा. अधिष्ठाता यांच्या आदेशान्वये / प्रशासकीय उपलब्धतेनुसार कमी किंवा वाढवली जाऊ शकते.
१४. पुरवठा धारकाचे पुर्ण नाव, भ्रमणध्वनी आणि संपुर्ण पत्ता, संकेतांक क्र. (Email address along with Website) हा Envelope वर नमुद असणे आवश्यक आहे.



अधिष्ठाता,
शासकीय वैद्यकीय महाविद्यालय,
अलिबाग रायगड
अधिष्ठाता,
शासकीय वैद्यकीय महाविद्यालय,
अलिबाग-रायगड

प्रत - शासकीय वैद्यकीय महाविद्यालय, तसेच जिल्हा सामान्य रुग्णालय, अलिबाग-रायगड यांच्या सुचना फलकावर दरपात्रक मागविण्या संदर्भात दिनांक :- ०८/०२/२०२४ सुचना देण्यात आल्या आहेत.

ANNEXURE- A

Demonstration Letter

To,
GOVERNMENT MEDICAL COLLEGE, ALIBAG

Ref: Bid Number

Our Engineer has Given the Demonstration of _____ product
Mentioned in Bid and We Ready to Complete the Supply, Installation and
Training Work.

Sign and Stamp of Bidder

Demonstration Satisfactory Report - Yes / No

Signature & Stamp of Concerned HOD

ANNEXURE - B
Technical Compliance Chart
 (On Bidders Letter Head)

S r N o	Bid specifications as asked in the Bid form	Equivalent specifications quoted by the Bided with the name of the manufacturer	Whether the Bid quoted by the supplier is as per specifications asked for. Indicate against each item clearly YES / NO.	If the Bid quoted is not as per specifications, then variations/ deviations should be clearly indicated against each item asked for.
1	2	3	4	5