

		
<b>महाराष्ट्र शासन</b> <b>शासकीय वैद्यकीय महाविद्यालय, अलिबाग, जि. रायगड.</b> जिल्हा सामान्य रुग्णालय परिसर, बीच रोड, लिमये वाडी, अलिबाग, जि. रायगड - ४०२ २०१ <b>GOVERNMENT MEDICAL COLLEGE, ALIBAG - RAIGAD</b> DISTRICT CIVIL HOSPITAL AREA, ALIBAG BEACH ROAD, LIMYEWADI, ALIBAG DIST. RAIGAD - 402 201		
Email- deangmcalibag@gmail.com	Phone No. ०२१४१-२९९२१४	www.gmchalibag.in/gmch
जा.क्र. शावैमअ/ Pathology Reagents / दरपत्रक मागणी/ 2094 / २०२५	दिनांक. ०९.०७.२०२५	

दरपत्रक मागणी पत्र.

प्रति,

-----  
-----  
-----

विषय :- शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, अलिबाग-रायगड या संस्थेच्या विकृतीशास्त्र विभागा करीता आवश्यक Reagents ची खरेदी करण्याबाबत...

अलिबाग येथे १०० विद्यार्थी प्रवेश क्षमतेचे शासकीय वैद्यकीय महाविद्यालय सन २०२१-२२ या शैक्षणिक वर्षापासून सुरू करण्यास शासनाने व राष्ट्रीय आयुर्विज्ञान आयोग, नवी दिल्ली, यांनी मान्यता प्रदान केलेली आहे. या संस्थेतील विकृतीशास्त्र विभागातील Hematology, Histopathology Lab, Cytopathology च्या यंत्राकरिता सद्यस्थितीत विभागात उपलब्ध असलेले Reagents अपूरे पडत असल्यामुळे अतिरिक्त Reagents ची तातडीची आवश्यकता आहे. त्याअनुषंगाने त्यासंबंधित बाबी ची खरेदी ही खालील नमुन्यात मोहरबंद लिफाफा विहित मुदतीत या कार्यालयाच्या आवक शाखेत खालील अटी व शर्ती नुसार सादर करावेत.

Sr. no	Drug/ Reagent name	Specifications	Pack size (Unit)	Price Per item Including Taxes
1	Liquid Handwash	Liquid hand wash containing Triclosan, Benzalkoniumchloride and moisturizer with fragrance colourq.s., owners quality control certificate. Plastic bottle with spring pourer 500ml	1x500ml	
2	Spirit	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified Plastic bottle with spring pourer 500ml	1x500ml	
3	Leishman stain	For staining slides, owners quality control certificate, CE notified / conformed to WHO-GMP	1x250 ml	
4	Absolute alcohol	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. CAS Number:67-63-0 Histological grade for histotechnique.	1x500ml	
5	Immersion oil	Pure grade D30 For Microscopy, Owners quality control certificate	1x25ml	
6	Methanol 500ml	A.R quality owners quality control certificate, pure form, more than 95%	1x500ml	



Sr. no	Drug/ Reagent name	Specifications	Pack size (Unit)	Price Per item Including Taxes
7	APTT reagent	Activated partial thromboplastin time, lyophilized with solvent / liquid stable, GMP Schedule M certificate product wise and testing facilities wise. CE notified 1 ml	1x2 ml	
8	Thromboplastin	Used for clotting time, Heparin insensitive upto 1.5 IU/ml, lyophilized with solvent / liquid stable, ISI 0.9 to 1.1, GMP Schedule M certificate product wise and testing facilities wise. One ml each (Abnormal and Normal) control should be provided free with every kit separately (120 ml).	1x5ml	
9	Malaria antigen rapid test	Rapid Malaria Antigen test for detection of Pf & Pv	1x25	
10	Microscope lens cleaner	CE notified / conformed to WHO-GMP	1x150 ml	
11	Gloves latex, Size 6	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485	100 pairs	
12	Gloves latex Size 6.5	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485	100 pairs	
13	Gloves latex Size 7	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485	100 pairs	
14	Gauze 90cmx18m	Material: Absorbent gauze, 100% cotton. Width: 65 to 90cm.Length:100m.Single use, Non-sterile.	1 roll	
15	Adhesive tape 2.5cmx5m	Adhesive tape cloth based with porous and hypoallergenic adhesive mass Having Uniform Spiked Impression size 2.5cm.x 5mtr	1 roll	
16	Anti A	(Monoclonal).Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website-cdco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified	1x10ml	
17	Anti B	(Monoclonal).Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website-cdco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified	1x10ml	
18	Anti D	(Monoclonal).Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website-cdco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified	1x10ml	
19	Cotton roll	Absorbent cotton wool (Nonsterile)	1x1kg	
20	Giemsa stain kit	Staining powder with buffer, owner quality certificate, CE notified / conformed to WHO-GMP	1x500ml	
21	Slide cabinet 50000 slide capacity	1. For keeping 75 x 25 mm glass slides in vertical position one after the other 2. Completely made of CRC, S.S 304 medical grade steel including slide carrying drawers and with powder coated paint finish 3. The cabinet should have multiple drawers to keep slides in vertical position 4. Each drawer should move in a slot and should be completely removable for easy lifting of slides 5. Can store minimum 50000 slides capacity 6. The doors should be lockable and to provide lock and key 7. Should be antirust, antifungal 8. Should have wheels for easy shifting/ movement	1 unit	
22	Funnel with long stem 150mm	ISO/CE notified, Borosilicate or equivalent glass 150mm	1 unit	



Sr. no	Drug/ Reagent name	Specifications	Pack size (Unit)	Price Per item Including Taxes
23	Apron plastic disposable	Non boilable, light weight, Thickness not less than 50 micron, green colour, water proof, tapes at neck & waist, length upto knee (length-48 inches chest 34 inches). Individually packed	1x10	
24	Wash bottle 500ml	Plastic with pourer wash bottle 500 ml	1x500ml	
25	Whatman filter paper	400x470mm, 68x58 cm equivalent to grade 1 imported, owners quality certified, CE notified	1 x 100 pcs	
26	Drape disposable plain sheet 100x150cm	Drape disposable poly plain Sheet 100cm x 150cm	1x10	
27	Acetone	A.R quality ,ISO certified, WHO/GMP ed to and CE notified	1x500 ml	
28	Diamond pencil	For marking glass slide, pointed metal tip, often made of diamond, handle is typically made of wood, providing a comfortable grip for precise writing	1 unit	
29	Glass petridish	120x25 mm, borosilicate glass, autoclavable, ISO certified, WHO/GMP confirmed to and CE notified	1 unit	
30	Litmus paper	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x20	
31	Reagent bottle with flat stopper	Reagent bottle with flat stopper 500ml, ISO/CE notified, Borosilicate or equivalent glass	1x500ml	
32	Compostable black garbage bag	Compostable Black Garbage Bag Size 18x22" Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."	1x100	
33	Compostable red garbage bag	Compostable Red Garbage Bag Size 18x22" Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."	1x100	
34	Compostable yellow garbage bag	Compostable Yellow Garbage Bag Size 18 X 22" Thickness: 70 micron Quality: Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."	1x100	
35	Plastic container	1000 ml plastic round bottle with lid, neck size 8 cm to 10 cm, Nonsterile	1 unit	
36	Plastic container	1500 ml plastic round bottle with lid, neck size 10 cm to 12cm, Nonsterile	1 unit	
37	Plastic container	2000 ml plastic round bottle with lid, neck size 12 cm to 15 cm, Nonsterile	1 unit	
38	Reticulocyte stain	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x250 ml	
39	Alcohol based surgical hand antiseptic Solution 500 ml bottle with dispenser	Active Ingredient in 100 gm- 2 propanol- 45.0 g , 1 Propanol 30.0 g Alcohol based surgical hand antiseptic containing macetronium ethyl sulphate 0.2 gm / 100 gm or 0.5 % to 1 % chlorhexidine gluconate with suitable emollient & moisturizer 500 ml Container To be accompanied with following valid test reports: 1) EN 1500 (HygenicHandrub) 2) EN 12791 (surgical Hand Disinfection) 3) EN 14348 ( Mycobactericidal activity) 4 ) EN 14476 (Virucidal activity) 5) ISO 10993-10 (skin Irritation Test	1x500ml	
40	Cytology fixative spray	Polyethylene glycol fixative with fine mist like spray dispenser to enable the rapid fixation of cytology cell spreads. ISO certified.	1x50ml	
41	Coplin jar	Material: Glass:10 slides capacity vertically	1 unit	
42	Glycerin	AR quality, owners quality control certificate, pure form, clear colorless liquid	1x 1000ml	
43	Rapid PAP staining kit	WHO GMP confirmed and should give certificate of analysis. Used for staining PAP smears	1x250 tests	
44	Sodium Hypochlorite	Sodium Hypochlorite Solution U.S.P.containing not less than 4% and not more than 6% by weight of Sodium Hypochlorite Solution. Required item wise WHO GMP 5 lit Can	1x5 lit	
45	Picric acid	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x 500 gm	



Sr. no	Drug/ Reagent name	Specifications	Pack size (Unit)	Price Per item Including Taxes
46	Long Knife	Must be of High quality Stainless Steel • Must have Long lasting edge, Must have Molded ergonomic handle with finger guard Blade Length of 14inches, sharp	1 unit	
47	Microtome blade – High profile	Stainless steel high profile,PTFE coated blade, disposable, 3-5 micron sections, 34°/80mm,Compatible with Rotary Microtome HM 355S	1x50	
48	Microtome blade – Low profile	Stainless steel high profile,PTFE coated blade, disposable, 35°/75mm, 3-5 micron sections for Rotary Microtome HM 355S	1x50	
49	Manual Weighing scale	Manual Platform Type Weighing Scale/Balance of 100gm-10Kg Capacity, ISI Marked	1 unit	
50	Measuring Scale	Stainless Steel 30 inch long	1 unit	
51	Staining jar 10cmx15cm glass rectangular (25 slides)	Colour: White, transparent Material: Glass; Closure Type Glass lid glass rectangular 10cmx 15cmHolding capacity (25 slides in stainless steel rack)	1 unit	
52	Blotting Paper	140x90mm Medium fibre free	1x100 sheet	
53	Tissue marking dye	Perfect marking of the margins, for fresh and fixed tissue, water and alcohol resistance of the marking in the processing and staining phases, Kit consists of 3 colour black, blue yellow and red	4x50ml	
54	Glycerol	98-100%, colorless, A.R quality, owners quality control certificate,	1x1 L	
55	Potassium Nitrate	A.R quality, owners quality control certificate,	1x100g	
56	Potassium Acetate	A.R quality, owners quality control certificate,	1x100g	
58	Sodium dihydrogen phosphate, monohydrate (NaH <sub>2</sub> PO <sub>4</sub> ·H <sub>2</sub> O)	White semi-transparent, granules or powder or solid, water soluble, FDA/GMP confirmed to/ Owners quality control certificate	1x50g	
59	Disodium hydrogen phosphate, anhydrous (Na <sub>2</sub> HPO <sub>4</sub> )	White to off-white granules or powder or solid, water soluble, FDA/GMP confirmed to/ Owners quality control certificate	1x30g	
60	Hematoxyline Delafield staining solution	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides	1x500ml	
61	Eosin stain solution	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides	1x500ml	
62	Isopropyl alcohol	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. CAS Number:67-63-0 Histological grade for histotechnique.	1x500ml	
63	Coverslip 22mm x 22mm	22X 22mm ISO certified OR confirmed to / CE notified	1x10gm	
64	Coverslip 22mm x 60mm	22 x 60 mm, ISO certified OR confirmed to / CE notified	1x10gm	
65	Stainless steel microscope slide tray	Stainless steel, 20 slide (26x76mm) capacity tray, rectangle shape	1 unit	
66	Rubber Mackintosh sheet	High quality, soft, blue red double sided, reuseable water proof, rubber mackintosh sheet, High operating temperature, Fair abrasion resistance, Good chemical resistance, Material: Rubber, Temperature: -60°C to 250°C	1x10mtr	
67	Plastic droppers	Disposable plastic dropper 0.2 ml	1x500 pcs	
68	Metal stirrers	Metal stirrers (Cat no. 804xF000100) for hemostar XF 2.0 , haemostasis analyzer.	1x100pcs	
69	Cuvettes	Plastic cuvettes (Cat no. 804XF000100) for hemostar XF 2.0, haemostasis analyzer.	1x100pcs	
70	Tissue Embedding Mold 2	Stainless steel molds 15 x 15 x 5 mm	1x 12	
71	Tissue Embedding Mold 3	Stainless steel mold 24 x 24 x 5 mm	1x 12	
72	Steel Tissue cassette	Stainless steel tissue cassettes with lid, rectangular, reusable (Medium)	1 Unit	



Sr. no	Drug/ Reagent name	Specifications	Pack size (Unit)	Price Per item Including Taxes
73	Plastic Tissue cassette	Plastic tissue cassettes with separate lid, rectangular (Medium)	1x100	
74	Plastic Tissue embedding ring	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides	25 X 25 X 12 mm	
75	Toluidine blue 1%	1) It should be a tissue staining solution of toluidine blue 1% dye. It should be available in a bottle with 500 ml of toluidine blue 1% 2) Easy-to-use dispenser prevents dye spillage or drying out it should have a property that it can be added directly to your tissue or on the processor. 3) it should not interfere with routine H&E, special stains, or IHC Formalin-based dyewill not evaporate as rapidly as similar alcohol-based alternatives. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x500 ml	
76	Glass Beaker	ISO/CE notified, Borosilicate glass	1x500 ml	
77	Measuring glass cylinder	ISO/CE notified, Borosilicate glass	1x500 ml	
78	Measuring glass cylinder	ISO/CE notified, Borosilicate glass	1x1000 ml	
79	Myeloperoxidase stain	For leukemia diagnosis. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x10 tests	
80	Periodic acid Schiff stain	For fungal and storage diseases diagnosis. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x500 ml	
81	Reticulin stain	For bone marrow biopsy staining. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x100ml	
82	Prussian blue stain	For iron stores staining. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	1x50 gm	
83	Slide labels	25x13 mm, white color, solvent and xylene resistant	1x100 labels /pack	
84	Acid fast bacilli (AFB) kit	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified	2x100ml	

अटी व शर्ती.

१. दरपत्रक मोहरबंद लिफाफा , दि.१६/०७/२०२५ रोजी साय ०५:०० वाजेपर्यंत कार्यालयास पोहचतील या बेताने पाठवावे. जाहीर केलेल्या तारखेनंतर अथवा कार्यालयीन वेळेनंतर आलेले दरपत्रक विचारात घेतले जाणार नाही/ग्राह्य धरले जाणार नाही.
२. दरपत्रक मोहोरबंद लिफाफा अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग. जि.रायगड ४०२२०१ (जिल्हा रुग्णालय आवार) यांचे नावे विहित मुदतीत कार्यालयास सादर करावा. त्याचप्रमाणे दरपत्रक मोहरबंद लिफाफ्यावर दरपत्रक मागणी क्रमांक, तारीख, व विभागाचे नाव **Pathology Reagents** असे लिहून सादर करावे.
३. लिफाफ्यावर पुरवठाधारकाचे पुर्ण नाव, भ्रमणध्वनी क्र. आणि संपूर्ण पत्ता, संकेतांक क्र. (Email address) लिफाफा जमा करण्याचा अंतिम दिनांक (१६/०७/२०२५), दरपत्रक धारकाच्या नावाचा शिक्का इ. नमुद असणे आवश्यक आहे.
४. दर पत्रकात नमुद केलेले दर सन २०२५-२६ या वित्तीय वर्षासाठी ग्राह्य धरले जातील.
५. प्रस्तावित उद्दिष्टांकरिता दरपत्रक आकारण्यात आलेले दर हे बाजारभावापेक्षा व इ-औषधी वैद्यकीय शिक्षण आणि संशोधन संचालनालय महाराष्ट्र संकेतस्थळावर दिलेले दरांपेक्षा (<https://medd-eproc.in/IMCS/hissso/>) अधिक नसावेत.
६. सर्वात कमी (L-१) दरपत्रकास खरेदी समितीच्या मान्यता अन्वये पुरवठा आदेशाकरिता विचार करण्यात येईल.
७. सदर उद्दिष्टांची किंमत, सर्व करांसह (including all taxes) एकुण किंमत अशी दर्शविण्यात यावी.



८. दरपत्रकात खाडाखोड करण्यात येवू नये. दरपत्रक हे पुर्णपणे छापील असावे. हस्तलिखित दरपत्रके ग्राह्य धरले जाणार नाहीत.
९. दरपत्रक मोहोरबंद लिफाफ्यात सादर करावयाचे दस्ताऐवज ( Documents Required)
  - १) अधिकृत लेटरहेडवर दरपत्रक दिलेल्या विनिर्दिष्टेनुसार असावेत २).जी.एस.टी.क्रमांक पत्र.
  - ३) दुकानाचा परवाना क्रमांक ४) Prebid demonstration report सोबत जोडलेल्या नमुन्यानुसार सदर दस्ताऐवज सादर केले नाही व अपुर्ण (विना स्वाक्षरी) असल्यास, तर सदरील पुरवठाधारकाचे दर न्यूनतम असले तरीही दरपत्रक ग्राह्य धरले जाणार नाही.
१०. Pre-bid demonstration करीता आणलेल्या व दाखविलेल्या Reagents यांचाच पुरवठा, पुरवठा आदेश मिळाल्यानंतर करणे बंधनकारक असेल. नमुने व पुरवठा यामध्ये तफावत आढळल्यास पुरवठा स्विकारले जाणार नाही व संबंधित पुरवठा दारास दोन वर्षा करिता वर्ज्य (Black Listed)सुचित केले जाईल.
११. Prebid demonstration करीता पुरवठादाराने स्वखर्चाने या कार्यालयात कार्यालयीन वेळेत व विहीत मुदतीत सादर करावे.
१२. ज्या पात्र दरपत्रक धारकाचे दर हे न्यूनतम असतील व दस्ताऐवज पुर्ण असतील अशा पुरवठादारास पुरवठा आदेश मिळाल्यानंतर दरपत्रकास आवश्यकते नुसार विकृतीशास्त्र विभागासाठीचा पुरवठा करण्याचे आदेश देण्यात येतील.
१३. पुरवठा आदेश मिळाल्यानंतर पंधरा दिवसांच्या आत पुरवठा न केल्यास पुरवठा आदेश रद्द करण्यात येईल, व द्वितीय न्यूनतम पात्र पुरवठाधारकाला प्रथम न्यूनतम दराने पुरवठा आदेश देण्यात येतील. पुरवठादाराने पुरवठा करण्यासाठी वाहतुक खर्च स्वतः करण्यात यावा.
१४. पुरवठा करण्यात येणारे **Pathology Reagents** चा कालबाह्य दिनांक (Expiry date) पुरवठा आदेश दिनांकापासून एक वर्षापेक्षा अधिक असावेत व त्यांचा दर्जा उत्तम असावा.
१५. संबंधित **Pathology Reagents** हे वैद्यकीय वापरासाठी योग्य असून, सर्व आवश्यक व गुणवत्ता निकष पूर्ण केलेले आहेत, याचे प्रमाणपत्र (Compliance Certificate) दरपत्रकासोबत सादर करणे अनिवार्य आहे.
१६. संस्थास्तरावरील औषधी, सर्जिकल आणि तदनुषंगीक बाबींची दरपत्रके, निविदा व स्थानिक खरेदी करतांना पुरवठादार/उत्पादक यांनी आवश्यकतेनुसार खालील बाबींची पूर्तता करणे बंधनकारक राहील.
  - Valid WHO GMP Certificate and WHO GMP product list or COPP for quoted Items.
  - In House test report for purchased item.
  - National Accreditation Board for Testing and Calibration Laboratories (NABL test report). Compulsory.
  - Non conviction certificate issued from concern FDA for manufacturer/Distributor.
  - For Consumables: ISO 13485 (International Organisation for Standardisation). 150 17025, 150 45001, ISO 14001. GMP (Good Manufacturing Practices)/Schedule M. Quality Management System (QMS) for Medical Devices, Central Drugs Standard Control Organisation (CDSCO) approved MD License.
१७. अपवादात्मक परिस्थितीमुळे उक्त कोणत्याही अटीत आवश्यकतेनुसार अंशतः बदल करणे अथवा प्राप्त झालेली दरपत्रके कोणतेही कारण न देता स्वीकारणे/नाकारणे तसेच कोणतेही स्तरावर खरेदी प्रक्रिया रद्द करणे हे अधिकार संस्थाप्रमुख (अधिष्ठाता) यांनी राखून ठेवले आहेत.



१८. पुरवठा पुर्ण झाल्यानंतर तीन प्रतीत देयके (Tax Invoice) हि अधिष्ठाता, शासकीय वैद्यकीय महाविद्यालय, अलिबाग जि. रायगड यांचे नावाने देण्यात यावेत.
१९. देयकाची अदायगी ही शासनाकडून अनुदान प्राप्त होण्याच्या अधीन राहून करण्यात येईल. देयक अदा करण्यास विलंब झाल्यास त्यावर कोणत्याही प्रकारचे विलंब व दंडात्मक शुल्क अदा केले जाणार नाही.



अधिष्ठाता,

शासकीय वैद्यकीय महाविद्यालय

अलिबाग रायगड.

शासकीय वैद्यकीय महाविद्यालय,

अलिबाग-रायगड

प्रत - शासकीय वैद्यकीय महाविद्यालय, तसेच जिल्हा सामान्य रुग्णालय, अलिबाग-रायगड यांच्या सुचना फलकावर  
दरपत्रक मागविण्या संदर्भात दिनांक : ०९/०७/२०२५ सुचना देण्यात आल्या आहेत.



Government Medical College Alibag  
Pre-Bid Demonstration Report (Format)

Sr. No	Product	Specification	Deviation from specification  Yes/ No	Name of the manufacturing company of the Reagent which supplier assures to provide for demonstration and for final supply. (to be filled by supplier)	Expiry date more than 92 months  Yes/ No	Demonstration is Satisfactory Yes/ No  (This column to be filled by Pathology department)
१	Liquid Handwash	Liquid hand wash containing Triclosan, Benzalkoniumchloride and moisturizer with fragrance colourq.s.owners quality control certificate. Plastic bottle with spring pourer 500ml				
२	Spirit	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified Plastic bottle with spring pourer 500ml				
३	Leishman stain	For staining slides, owners quality control certificate, CE notified / conformed to WHO-GMP				
४	Absolute alcohol	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. CAS Number:67-63-0 Histological grade for histotechnique.				
५	Immersion oil	Pure grade D30 For Microscopy, Owners quality control certificate				
६	Methanol 500ml	A.R quality owners quality control certificate, pure form, more than 95%				
७	APTT reagent	Activated partial thromboplastin time, lyophilized with solvent / liquid stable, GMP Schedule M certificate product wise and testing facilities wise. CE notified 1 ml				
८	Thromboplastin	Used for clotting time, Heparin insensitive upto 1.5 IU/ml, lyophilized with solvent / liquid stable, ISI 0.9 to 1.1, GMP Schedule M certificate product wise and testing facilities wise. One ml each (Abnormal and Normal) control should be provided free with every kit separately (120 ml).				
९	Malaria antigen rapid test	Rapid Malaria Antigen test for detection of Pf & Pv				
१०	Microscope lens cleaner	CE notified / conformed to WHO-GMP				
११	Gloves latex, Size 6	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485				
१२	Gloves latex Size 6.5	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485				



୧୩	Gloves latex Size 7	Disposable examination gloves made of soft latex, powder -free, highly elastic and tear-resistant, excellent fit, EN 455 1-4, CE certified, ISO 13485				
୧୪	Gauze 90cmx18m	Material: Absorbent gauze, 100% cotton. Width: 65 to 90cm.Length:100m.Single use, Non-sterile.				
୧୫	Adhesive tape 2.5cmx5m	Adhesive tape cloth based with porous and hypoallergenic adhesive mass Having Uniform Spiked Impression size 2.5cm.x 5mtr				
୧୬	Anti A	(Monoclonal),Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website- edco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified				
୧୭	Anti B	(Monoclonal),Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website- edco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified				
୧୮	Anti D	(Monoclonal),Measured volumes -10 ml, Absense of Turbidity particles. Red Panel should correspond to tested antibody with no reaction. Titre 1/256 Avidity less than 5 seconds. The manufacturer should be continuously be on list of Blood Bank Website- edco.nic.in, GMP Schedule M certificate product wise and testing facilities wise. FDA, CDSCO approved for blood bank 5 ml. ISO certified				
୧୯	Cotton roll	Absorbent cotton wool (Nonsterile)				
୨୦	Giemsa stain kit	Staining powder with buffer, owner quality certificate, CE notified / conformed to WHO-GMP				
୨୧	Slide cabinet 50000 slide capacity	1. For keeping 75 x 25 mm glass slides in vertical position one after the other 2. Completely made of CRC, S.S 304 medical grade steel including slide carrying drawers and with powder coated paint finish 3. The cabinet should have multiple drawers to keep slides in vertical position 4. Each drawer should move in a slot and should be completely removable for easy lifting of slides 5. Can store minimum 50000 slides capacity 6. The doors should be lockable and to provide lock and key 7. Should be antirust, antifungal 8. Should have wheels for easy shifting/ movement				



२२	Funnel with long stem 150mm	ISO/CE notified, Borosilicate or equivalent glass 150mm				
२३	Apron plastic disposable	Non boilable, light weight, Thickness not less than 50 micron, green colour, water proof, tapes at neck & waist, length upto knee (length-48 inches chest 34 inches). Individually packed				
२४	Wash bottle 500ml	Plastic with pourer wash bottle 500 ml				
२५	Whatman filter paper	400x470mm, 68x58 cm equivalent to grade 1 imported, owners quality certified, CE notified				
२६	Drape disposable plain sheet 100x150cm	Drape disposable poly plain Sheet 100cm x 150cm				
२७	Acetone	A.R quality ,ISO certified, WHO/GMP ed to and CE notified				
२८	Diamond pencil	For marking glass slide, pointed metal tip, often made of diamond, handle is typically made of wood, providing a comfortable grip for precise writing				
२९	Glass petridish	120x25 mm, borosilicate glass, autoclavable, ISO certified, WHO/GMP confirmed to and CE notified				
३०	Litmus paper	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
३१	Reagent bottle with flat stopper	Reagent bottle with flat stopper 500ml, ISO/CE notified, Borosilicate or equivalent glass				
३२	Compostable black garbage bag	Compostable Black Garbage Bag Size 18x22" Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."				
३३	Compostable red garbage bag	Compostable Red Garbage Bag Size 18x22" Thickness : 70 micron Quality : Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."				
34	Compostable yellow garbage bag	Compostable Yellow Garbage Bag Size 18 X 22" Thickness: 70 micron Quality: Double sealing chlorine free leak proof and complaint with Government of Maharashtra notifications of plastics."				
35	Plastic container	1000 ml plastic round bottle with lid, neck size 8 cm to 10 cm, Nonsterile				
36	Plastic container	1500 ml plastic round bottle with lid, neck size 10 cm to 12cm, Nonsterile				
37	Plastic container	2000 ml plastic round bottle with lid, neck size 12 cm to 15 cm, Nonsterile				
38	Reticulocyte stain	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
39	Alcohol based surgical hand antiseptic Solution 500 ml bottle with dispenser	Active Ingredient in 100 gm- 2 propanol- 45.0 g , 1 Propanol 30.0 g Alcohol based surgical hand antiseptic containing macetronium ethyl sulphate 0.2 gm / 100 gm or 0.5 % to 1 % chlorhexidine gluconate with suitable emollient & moisturizer 500 ml Container To be accompanied with following valid test reports: 1) EN 1500 (HygenicHandrub) 2) EN 12791 (surgical Hand Disinfection) 3) EN				



		14348 (Mycobactericidal activity) 4 ) EN 14476 (Virucidal activity) 5) ISO 10993-10 (skin Irritation Test)			
40	Cytology fixative spray	Polyethylene glycol fixative with fine mist like spray dispenser to enable the rapid fixation of cytology cell spreads. ISO certified.			
41	Coplin jar	Material: Glass;10 slides capacity vertically			
42	Glycerin	AR quality, owners quality control certificate, pure form, clear colorless liquid			
43	Rapid PAP staining kit	WHO GMP confirmed and should give certificate of analysis. Used for staining PAP smears			
44	Sodium Hypochlorite	Sodium Hypochlorite Solution U.S.P.containing not less than 4% and not more than 6% by weight of Sodium Hypochlorite Solution. Required item wise WHO GMP 5 lit Can			
45	Picric acid	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified			
46	Long Knife	Must be of High quality Stainless Steel • Must have Long lasting edge, Must have Molded ergonomic handle with finger guard Blade Length of 14inches, sharp			
47	Microtome blade – High profile	Stainless steel high profile,PTFE coated blade, disposable, 3-5 micron sections, 34°/80mm.Compatible with Rotary Microtome HM 355S			
48	Microtome blade – Low profile	Stainless steel high profile,PTFE coated blade, disposable, 35°/75mm. 3-5 micron sections for Rotary Microtome HM 355S			
49	Manual Weighing scale	Manual Platform Type Weighing Scale/Balance of 100gm-10Kg Capacity, ISI Marked			
50	Measuring Scale	Stainless Steel 30 inch long			
51	Staining jar 10cmx15cm glass rectangular (25 slides)	Colour: White, transparent Material: Glass; Closure Type Glass lid glass rectangular 10cmx 15cmHolding capacity (25 slides in stainless steel rack)			
52	Blotting Paper	140x90mm Medium fibre free			
53	Tissue marking dye	Perfect marking of the margins, for fresh and fixed tissue, water and alcohol resistance of the marking in the processing and staining phases. Kit consists of 3 colour black, blue yellow and red			
54	Glycerol	98-100%, colorless, A.R quality, owners quality control certificate,			
55	Potassium Nitrate	A.R quality, owners quality control certificate,			
56	Potassium Acetate	A.R quality, owners quality control certificate,			
58	Sodium dihydrogen phosphate, monohydrate (NaH <sub>2</sub> PO <sub>4</sub> ·H <sub>2</sub> O)	White semi-transparent, granules or powder or solid, water soluble, FDA/GMP confirmed to/ Owners quality control certificate			



59	Disodium hydrogen phosphate, anhydrous (Na <sub>2</sub> HPO <sub>4</sub> )	White to off-white granules or powder or solid, water soluble, FDA/GMP confirmed to/ Owners quality control certificate				
60	Hematoxyline Delafield staining solution	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides				
61	Eosin stain solution	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides				
62	Isopropyl alcohol	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. CAS Number:67-63-0 Histological grade for histotechnique.				
63	Coverslip 22mm x 22mm	22X 22mm ISO certified OR confirmed to / CE notified				
64	Coverslip 22mm x 60mm	22 x 60 mm, ISO certified OR confirmed to / CE notified				
65	Stainless steel microscope slide tray	Stainless steel, 20 slide (26x76mm) capacity tray, rectangle shape				
66	Rubber Mackintosh sheet	High quality, soft, blue red double sided, reuseable water proof, rubber mackintosh sheet, High operating temperature, Fair abrasion resistance, Good chemical resistance, Material: Rubber, Temperature: -60°C to 250°C				
67	Plastic droppers	Disposable plastic dropper 0.2 ml				
68	Metal stirrers	Metal stirrers (Cat no. 804xF000100) for hemostar XF 2.0 , haemostasis analyzer.				
69	Cuvettes	Plastic cuvettes (Cat no. 804XF000100) for hemostar XF 2.0, haemostasis analyzer.				
70	Tissue Embedding Mold 2	Stainless steel molds 15 x 15 x 5 mm				
71	Tissue Embedding Mold 3	Stainless steel mold 24 x 24 x 5 mm				
72	Steel Tissue cassette	Stainless steel tissue cassettes with lid, rectangular, reusable (Medium)				
73	Plastic Tissue cassette	Plastic tissue cassettes with separate lid, rectangular (Medium)				
74	Plastic Tissue embedding ring	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified. Used for staining histopathology slides				
75	Toluidine blue 1%	1) It should be a tissue staining solution of toluidine blue 1% dye. It should be available in a bottle with 500 ml of toluidine blue 1% 2) Easy-to-use dispenser prevents dye spillage or drying out it should have a property that it can be added directly to your tissue or on the processor. 3) it should not interfere with routine H&E, special stains, or IHC Formalin-based dyewill not evaporate as rapidly as similar alcohol-based alternatives. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
76	Glass Beaker	ISO/CE notified, Borosilicate glass				
77	Measuring glass cylinder	ISO/CE notified, Borosilicate glass				



78	Measuring glass cylinder	ISO/CE notified, Borosilicate glass				
79	Myeloperoxidase stain	For leukemia diagnosis. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
80	Periodic acid Schiff Stain	For fungal and storage diseases diagnosis. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
81	Reticulin stain	For bone marrow biopsy staining. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
82	Prussian blue stain	For iron stores staining. AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				
83	Slide labels	25x13 mm, white color, solvent and xylene resistant				
84	Acid fast bacilli (AFB) kit	AR quality and equivalent, ISO certified, WHO/GMP confirmed to and CE notified				

**Undertaking of the Supplier: I hereby confirm to supply above mentioned items/reagents in good condition and as per given specification and mentioned manufacturer with expiry date more than 12 months from the date of supply.**

Name of the Supplier:

Sign with Stamp:

----- For office use only -----

Demonstration of above mentioned items/ reagents was conducted at GMC Alibag.

Professor and Head  
Department of Pathology  
GMC Alibag